

## **Declaration of Conformity**

Quality Systems and Regulatory Compliance

Prestan is a developer and manufacturer of emergency medical training products. We employ a comprehensive set of processes and methods to insure that our products are safe, reliable and effective. Our quality systems require that our internal processes as well as those of our suppliers and contract manufacturers include statistically based inspection plans and audits to insure product consistency, compliance and control of non-conforming goods.

We certify that our products conform to all mandatory industry regulations and as well as a number of additional standards on a voluntary basis. The following listing identifies those standards which are adhered to subject to the standard's relevancy to each product, process and resource utilized.

0	FDA GMP, Class I	FDA's Good Manufacturing Practices: Quality System Requirements
0	ANSI/ASQC Z1.4	Statistical Sampling Procedures and Tables for Inspection of Materials and
		Finished Goods
0	RoHS (2011/65/EU + 2015/863/EU)	Restriction of Hazardous Substances in Electronic Products and Components
0	Toxics In Packaging Act	(and EU 94/62/EC) Restriction of Hazardous Substances in Packaging
0	UL94HB	Flammability Standard for Plastic Resins
0	EN 61326 / IEC 61326	Electro-Magnetic Compatibility (EMC): Radiated Emissions, Electrostatic
		Discharge and Radiated RF Susceptibility
0	IEC 61010-1	Safety Requirements for Electrical Equipment
0	IEC 60529	Degrees of Protection Provided by Device Enclosures (IP Code)
0	IPC-A-610-G	Acceptability of Electronic Assemblies: Class 2 Minimum
0	IPC-D-325A	Documentation Requirements for Printed Circuit Boards and Assemblies
0	IPC-4552A	Specification for Electroless Nickel/Immersion Gold (ENIG) Plating for Printed
		Circuit Boards
0	IPC-A-620B	Acceptance Requirements or Cable and Wire Harness Assemblies
0	IPC-7711B/7721B	Rework, Repair and Modification of Electronic Assemblies
0	SDS	Safety Data Sheets: Applicable Documents Maintained on File
0	ISTA 1A	Packaging Impact: Drop Test Procedure ISTA 1A

Name and Title of Signatory:

Chris M. Charlton, Vice President – Product Development Prestan Products, LLC